

# Toxo IgM

## IgM antibodies to Toxoplasma gondii (ELISA)

REF: DS180602



### Intended use

The DiaSino Toxo IgM assay is an enzyme-linked immunosorbent assay (ELISA) for the in vitro qualitative determination of IgM antibodies to Toxoplasma gondii in Human serum and plasma.

### Summary<sup>1-7</sup>

Toxoplasmosis is a parasitic disease caused by Toxoplasma gondii. Infections with Toxoplasmosis usually cause no obvious symptoms in adults. Occasionally, people may have a few weeks or months of mild, flu-like illness such as muscle aches and tender lymph nodes. In a small number of people, eye problems may develop. In those with a weak immune system, severe symptoms such as seizures and poor coordination may occur. If infected during pregnancy, a condition known as congenital Toxoplasmosis may affect the child. Up to half of the world's population is infected by Toxoplasmosis, but have no symptoms. In the United States, about 23% are affected and in some areas of the world this is up to 95%. About 200,000 cases of congenital Toxoplasmosis occur a year.

Diagnosis of Toxoplasmosis in humans is made by biological, serological, histological, or molecular methods. Serological testing can detect Toxo. gondii antibodies in blood serum, using methods including the Sabin-Feldman dye test (DT), the indirect hemagglutination assay, the indirect fluorescent antibody assay (IFA), the direct agglutination test, the latex agglutination test (LAT), the enzyme-linked immunosorbent assay (ELISA), and the immunosorbent agglutination assay test (IAAT).

The most commonly used tests to measure IgG antibody are the DT, the ELISA, the IFA, and the modified direct agglutination test. IgG antibodies usually appear within a week or two of infection, peak within one to two months, then decline at various rates. Toxoplasma IgG antibodies generally persist for life, and therefore may be present in the bloodstream as a result of either current or previous infection.

### Test principle

Capture method. Total duration of assay: **70 minutes**.

This kit uses capture ELISA principle to detect Toxo IgM. Polystyrene microwell strips are pre-coated with mouse anti-human IgM (anti-μ chain specific). During the first incubation step, all IgM antibodies binds to mouse anti-human IgM monoclonal antibody coated in microplate including the Anti-Toxo IgM, the wells are washed to remove unbound serum proteins, and Toxo antigens conjugated to the enzyme horseradish peroxidase (HRP-Conjugate) are added. During the second incubation step, if there is Toxo IgM antibody in sample, HRP-conjugated antigens will be bound to the anti-human Toxo IgM complexes previously formed and the unbound HRP-conjugate is then removed by washing. Substrate solution is then added and catalyzed by this complex, resulting in a chromogenic reaction. The resulting chromogenic reaction is measured as absorbance.

The amount of color intensity can be measured and it is proportional to the amount of antibody captured in the wells, and to the amount of antibody in the sample respectively.

### Reagents

#### Materials provided

- **Coated Microplate - symbol** TOXO IgM PLATE 8 x 12 strips, 96 wells.
- **Enzyme Conjugate - symbol** TOXO IgM CONJ 1 vial, 11.0 mL of HRP (horseradish peroxidase) labeled with Toxo IgM antigens. Contains 0.1% ProClin300 preservative.
- **Sample Diluent - symbol** TOXO IgM DILUT 1 vial, 11 mL, ready to use
- **Positive Control - symbol** TOXO IgM PC 1 vial, 0.2 mL.
- **Negative Control - symbol** TOXO IgM NC 1 vial, 0.2 mL
- **Substrate - symbol** SUBSTRATE 1 vial, 11mL, ready to use, (tetramethylbenzidine) TMB.
- **Stop Solution - symbol** STOP 1 vial, 6.0 mL of 1 mol/L sulfuric acid.
- **Wash Solution Concentrate - symbol** WASH 40X 1 vial, 25 mL (40X concentrated), PBS-Tween wash solution.
- **IFU - 1 copy.**
- **Plate Lid: 2 pieces.**

#### Materials required (but not provided)

- Microplate reader with 450nm and 620nm wavelength absorbent capability.
- Microplate washer.
- Incubator.
- Plate shaker.

# diasino

- Micropipettes and multichannel micropipettes delivering 50 µL with a precision of better than 1.5%.
- Absorbent paper.
- Distilled water

### Precautions and warnings

- For in vitro diagnostic use only.
- Exercise the normal precautions required for handling all laboratory reagents.
- Disposal of all waste material should be in accordance with local guidelines.
- Do not use reagents beyond the labeled expiry date.
- Do not mix or use components from kits with different batch codes.
- All the specimen and reaction wastes should be considered potentially biohazard. The handling of specimens and reaction wastes should be in accordance with the local regulations and guidelines.
- The Stop Solution contains sulfuric acid, which can cause severe burns. In the event of contact with eyes, rinse immediately with plenty of water and seek medical advice.
- Neutralized acids and other liquid waste should be decontaminated by adding a sufficient volume of sodium hypochlorite to obtain a final concentration of at least 1.0%. Exposure to 1.0% sodium hypochlorite for 30 minutes may be necessary to ensure effective decontamination.
- Some reagents contain 0.05% or 0.1% ProClin 300 which may cause sensitization by skin contact, which must therefore be avoided. Reagents and their containers must be disposed of safely. If swallowed, seek medical advice immediately and show this container or label.
- Substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
- For information on hazardous substances included in the kit please refer to the Materials Safety Data Sheet (MSDS), which is available on request.
- Do not smoke, drink, eat or apply cosmetics in the work area.
- Do not pipette by mouth. Wear protective clothing, disposable gloves and eye/face protection when handling samples and reagents. Wash hands after use.
- If any of the reagents comes into contact with the skin or eyes, wash the area extensively with water.

### Storage and stability

- Store at 2-8°C.
- Seal and return unused reagents to 2-8°C, under which conditions the stability will be retained for 2 months, or until the labeled expiry date, whichever is earlier.

### Specimen collection and preparation

- Human serum or plasma is recommended for this assay.
- Cap and store the samples at 18-25 °C for no more than 8 hours. Stable for 7 days at 2-8°C, and 1 month at -20°C. Freeze only once.
- Do not use heat-inactivated samples.
- Sediments and suspended solids in samples may interfere with the test result which should be removed by centrifugation. Ensure that complete clot formation in serum samples has taken place prior to centrifugation.
- Avoid grossly hemolytic, lipemic or turbid samples.

### Wash solution (40X dilution)

Add deionized water to the 40X concentrated Wash Solution Concentrate. Dilute 25 mL of Wash Solution Concentrate with 975 mL of deionized water to a final volume of 1000 mL. Stable for 2 months at room temperature.

### Test procedure

Ensure the patients' samples, calibrators, and controls are at ambient temperature (18-25 °C) before measurement. Mix all reagents through gently inverting prior to use.

- **Preparation:** Mark two wells as Negative control (e.g. B1, C1), two wells as Positive control (e.g. D1, E1) and one Blank (e.g. A1, neither samples nor HRP-Conjugate should be added into the Blank well). If the results will be determined by using dual wavelength plate reader, the requirement for use of Blank well could be omitted. Use only number of strips required for the test.
- **Adding Sample Diluent:** Add **100 µL** of Sample Diluent into their respective wells except the Blank.
- **Adding Sample:** Add **10 µL** of **Positive control, Negative control, and Specimen** into their respective wells except the Blank. Mix by tapping the plate gently. Use a separate disposal pipette tip for each specimen to avoid cross-contamination.

**Note:** After adding Sample, the reagents in wells turns **Blue** color from **Green**.

# Toxo IgM

## IgM antibodies to *Toxoplasma gondii* (ELISA)

- **Incubating:** Cover the plate with the plate cover and incubate for **30 minutes at 37°C**.
- **Washing:** At the end of the incubation, remove and discard the plate cover. Wash each well **5 times** with diluted Wash Buffer. Each time allow the microwells to soak for **30-60 seconds**. After the final washing cycle, turn down the plate onto blotting paper or clean towel and tap it to remove any remainders.
- **Adding Conjugate:** Add **100 µL of Conjugate** into each well except the Blank.
- **Incubating:** Cover the plate with the plate cover and incubate for **30 minutes at 37°C**.
- **Washing:** At the end of the incubation, remove and discard the plate cover. Wash each well **5 times** with diluted Wash Buffer. Each time allow the microwells to soak for **30-60 seconds**. After the final washing cycle, turn down the plate onto blotting paper or clean towel and tap it to remove any remainders.
- **Coloring:** Add **100 µL of Substrate** into each well including the Blank. Incubate the plate at room temperature for 10 minutes avoiding light. The enzymatic reaction between the TMB substrate and the HRP-Conjugate produces blue color in Positive control and in positive sample wells.
- **Stopping:** Using a multichannel pipette or manually, add **50 µL of Stop Solution** into each well and mix gently. Intensive yellow color develops in Positive control and Toxo IgM positive sample wells.
- **Measuring:** Calibrate the plate reader with the Blank well and read the absorbance at **450nm**. If a dual filter instrument is used, set the reference wavelength at **630nm**.

Calculate the Cut-off value and evaluate the results. (**Note:** read the absorbance within **10 minutes** after stopping the reaction).

### Quality control and calculation

- Read the sample's optical density (OD) at 450nm with a micro plate reader.
- Mean negative control OD value  $\leq 0.1$  and positive control OD value  $\geq 0.8$ , the test is valid, otherwise the test is invalid.
- Cut-Off value (C.O.) = Mean negative control OD value + 0.10 (Calculated by 0.05 when Mean negative control OD value is  $\leq 0.05$ , calculated by actual value when Mean negative control OD value is  $> 0.05$ ).

### Positive Results: Sample O.D value $\geq$ C.O.

Specimens giving an absorbance equal to or greater than the Cut-Off value are considered initially reactive, which indicates that Toxo IgM has probably been detected using Toxo IgM ELISA. All initially reactive specimens should be retested in duplicates using Toxo IgM ELISA before the final assay results interpretation. Repeatedly reactive specimens can be considered positive for Toxo IgM with Toxo IgM ELISA.

### Negative Results: Sample OD value $<$ C.O

Specimens giving absorbance less than the Cut-Off value are negative for this assay, which indicates that no Toxo IgM has been detected with Toxo IgM ELISA, therefore the patient is probably not infected with Toxo and the blood does not contain Toxo IgM.

### Limitations

- Positive results must be confirmed with another available method and interpreted in conjunction with the patient clinical information.
- Clinical diagnosis should not be established based on a single test result. It should integrate clinical and other laboratory data and findings.

### Performance Characteristics

#### Specificity

A study of 216 individuals was tested with DiaSino ELISA kits and found that the specificity was 99.05%.

#### Sensitivity

Among 181 Toxo IgM confirmed were positive when tested with this ELISA, and 180 samples of which were detected with positive, the sensitivity was 99.45%.

#### Precision

##### Intra-assay

CV  $\leq 15\%$

Within-run precision has been determined by using 15 replicates of three specimens: a low positive, medium positive and a high positive.

##### Inter-assay

CV  $\leq 20\%$

Between-run precision has been determined by 3 independent assays on the same three specimens: a low positive, medium positive and a high positive. Three different lots of the Toxo IgM ELISA Test Kit have been tested using these specimens over a 5-day period.











### Analytical specificity

- Interferences are not observed up to a concentration of 1 mg/mL Acetaminophen, 0.2 mg/mL Gentistic Acid, 0.1 mg/mL Ascorbic Acid, 0.1 mg/mL Acetosalisilic Acid, 0.1 mg/mL Caffeine, 0.6 mg/mL Oxalic Acid, 2 mg/mL Bilirubin, 2 mg/mL Hemoglobin, and 1% Ethanol.
- Rheumatoid factors do not interfere with the test.
- Cross-Reactivity are not observed in Toxo IgG, Syphilis, HBsAg, HCV, RF and HCG positive specimens.

### Reference

1. "Parasites – Toxoplasmosis (Toxoplasma infection) Disease". July 10, 2014. Archived from the original on 22 August 2015. Retrieved 22 August 2015.
2. Hunter, CA; Sibley, LD (November 2012). "Modulation of innate immunity by *Toxoplasma gondii* virulence effectors". *Nature Reviews Microbiology*. 10 (11): 766–78. doi:10.1038/nrmicro2858. PMC 3689224. PMID 23070557.
3. "Parasites – Toxoplasmosis (Toxoplasma infection) Epidemiology & Risk Factors". March 26, 2015. Archived from the original on 23 August 2015. Retrieved 22 August 2015.
4. Torgerson, Paul R; Mastroiacovo, Pierpaolo (2013). "The global burden of congenital Toxoplasmosis: a systematic review". *Bulletin of the World Health Organization*. 91 (7): 501–508. doi:10.2471/BLT.12.111732. ISSN 0042-9686. PMC 3699792. PMID 23825877.
5. Flegr J, Prandota J, Sovičková M, Israli ZH (March 2014). "Toxoplasmosis—a global threat. Correlation of latent Toxoplasmosis with specific disease burden in a set of 88 countries". *PLoS ONE*. 9 (3): e90203. Bibcode:2014PLoS...990203F. doi:10.1371/journal.pone.0090203. PMC 3963851. PMID 24662942. Toxoplasmosis is becoming a global health hazard as it infects 30–50% of the world human population.
6. Jones JL, Parise ME, Fiore AE (2014). "Neglected parasitic infections in the United States: Toxoplasmosis". *American Journal of Tropical Medicine and Hygiene*. 90 (5): 794–9. doi:10.4269/ajtmh.13-0722. PMC 4015566. PMID 24808246.
7. Montoya JG (2002). "Laboratory diagnosis of *Toxoplasma gondii* infection and Toxoplasmosis". *The Journal of Infectious Diseases*. 185 Suppl 1: S73–82. doi:10.1086/338827. PMID 11865443.

### Symbols

			
In vitro diagnostic medical device	Temperature limit	Consult instructions for use	Catalog number
			
Batch code	Date of manufacture	Use-by date	Contains sufficient for <n> tests
	 Do not use if package is damaged and consult instructions for use		
Manufacturer			



## diasino

DiaSino Laboratories Co., Ltd

No.68 Jingnansi Road National Eco & Tech Development Area  
Zhengzhou 450000, P.R. China. info@diasino.com