

FT3

Free Triiodothyronine (FIA)

REF: IN017704



25

Intended use

The infinosiTM Free T3 is a fluorescence immunoassay (FIA) for the in vitro quantitative determination of free triiodothyronine (FT3) in human serum or plasma. It is useful as an aid in management and monitoring of measurement in the assessment of thyroid function. For professional use only.

Summary

Triiodothyronine is one of the thyroid hormones present in serum which regulates metabolism. Determination of this hormone concentration is important for the diagnostic differentiation of euthyroid, hyperthyroid, and hypothyroid states. The major fraction of total triiodothyronine is bound to the transport proteins (TBG, prealbumin, albumin). Free triiodothyronine (FT3) is the physiologically active form of the thyroid hormone triiodothyronine (T3). The determination of free T3 has the advantage of being independent of changes in the concentrations and binding properties of the binding proteins; additional determination of a binding parameter (T-uptake, TBG) is therefore unnecessary.^{1,2,3}

In normal thyroid function, as the concentrations of the carrier proteins alter, the total T3 level changes so that the FT3 concentration remains constant.⁴ Thus, measurements of FT3 concentrations correlate more reliably with clinical status than total T3 levels. For example, the increase in total T3 levels associated with pregnancy, oral contraceptives and estrogen therapy result in higher total T3 levels while the FT3 concentration remains basically unchanged.⁵ In addition, it has been found that the mean FT3 value has a gradient decreasing from young to older.⁶

Test principle

Competitive principle. Total duration of assay: **15 minutes**

Sample is added to the sample well of the test, the fluorescence-labeled detector FT3 antibodies bind to FT3 antigens in blood specimen and form immune complexes. As the complexes migrate on the nitrocellulose matrix by capillary action, it can't be captured by FT3 antigens that have been immobilized on test strip, otherwise the excess unbound fluorescence-labeled detector FT3 antibodies are captured. Thus the more FT3 in blood, the less unbound fluorescence-labeled antibodies accumulated on test strip. Signal intensity of detector FT3 antibodies reflect the amount of antigens and are processed in the instrument for infinosiTM tests to determine the FT3 concentration in blood.

Reagents

Materials provided

- **Test Cartridge**, 25 pcs, individually packaged
- **ID Chip or QR code of Calibration Curve**, 1 pcs
- **Sample Buffer**, 1 x 2.0 mL
- **Detector Tube (lyophilized)**: 25 tubes **NOTE: Once the lid is opened, it must be used immediately**
- **IFU**, 1 copy

Materials required (but not provided)

- infinosiTM FIA analyzer
- FT3 control (DiaSino control is recommended)
- Specimen collection containers
- Transfer pipette set (100 µL size)
- Centrifuge (for plasma and serum only)
- Timer

Precautions and warnings

- For in vitro diagnostic use only.
- Carefully follow the instructions and procedures described in this instructions before testing.
- The test cartridge should remain in its original sealed pouch until ready to use. Do not use it if the pouch is damaged or the seal is broken.
- Do not use reagents beyond the labeled expiry date.
- Do not mix or use components from kits with different Lots.
- Don't use Test Cartridge if its Lot does not match with ID Chip that is inserted onto the instrument.
- The infinosiTM FT3 should be used only in conjunction with the instrument for infinosiTM tests.
- The tests should be applied by professionally trained staff working in certified laboratories at some remove from the patient and clinic at which the sample is taken by qualified medical personnel.
- infinosiTM FT3 assay is single use only. Do not reuse it.

infinosiTM

- The Test Cartridge and instrument for infinosiTM tests should be used away from vibration and magnetic field. During normal usage, the Test Cartridge may generate slight vibration, which should be regarded as normal.
- Use separate clean pipette tips and buffer tubes for different specimens. The pipette tips and detector buffer tubes should be used for one specimen only.
- Do not smoke, eat, or drink in areas in which specimens or kit reagents are handled.
- Blood specimens, used test cartridges, pipette tips and sample buffer tubes are potentially infectious. Proper laboratory safety techniques, handling and disposal methods should be followed in accordance with standard procedures and relevant regulations observed by microbiological hazard materials.
- The results should be interpreted by the physician along with clinical findings and other laboratory test results.

Incident report

Any suspected serious incidents related to this assay shall be immediately reported to DiaSino, DiaSino's Authorized Representative in the EU, and the national competent authorities of the Member States where the users and/or patients are located.

Storage and stability

- Store the test kit at 2-30°C, the stability is up to the expiration date printed on package.
- Store all the other components at 2-30°C, the stability is up to the expiration date printed on package.
- Test cartridge and sample buffer should be used within 1 hour after opening the pack.

Specimen collection and preparation

- The test can be performed with either serum or plasma.
- Collect serum samples in accordance with correct medical practices.
- Using standard phlebotomy procedure, collect a venipuncture whole blood specimen using a blood collection tube. If collecting plasma use a blood collection tube containing suitable anticoagulant (EDTA recommended).
- Separate the serum/plasma from blood as soon as possible to avoid hemolysis.
- Test should be performed immediately after the specimens have been collected. Do not leave the specimens at room temperature for prolonged periods. Specimens may be stored at 2-8°C for up to 3 days. For long-term storage, specimens should be kept below -20°C.

Quality control

- Quality control tests are a part of the good testing practice to confirm the expected results and validity of the assay and should be performed at regular intervals.
- The control tests should be performed immediately after opening a new test lot to ensure the test performance is not altered.
- Quality control tests should also be performed whenever there is any question concerning the validity of the test results.
- Control materials are provided on demand with infinosiTM tests. For more information regarding obtaining the control materials, contact [DiaSino Laboratories Co., Ltd](#) for assistance.

Test setup

- Ensure that the lot number of the cartridge matches that of the sample buffer, and the ID Chip.
- If the sealed cartridge and sample buffer have been stored in refrigerator, place them at room temperature (18-25 °C) at least 30 minutes before measurement.
- Turn on the instrument for infinosiTM tests.
Refer to the "instrument for infinosiTM tests Operation Manual" for the complete information and operating instructions.

Test procedure

1. Insert ID Chip into the instrument for infinosiTM tests or Scan the QR code to read the calibration curve.
2. Using a pipette to transfer **70 µL of Sample Buffer**, and **70 µL of sample (Human plasma/serum)** to the **Detector Tube (lyophilized)**. Close the lid of the tube and mix the sample thoroughly for 5-10 seconds by tapping or inverting the tube, **then incubate the sample mixture at room temperature for 5 minutes.**
3. Pipette out **100 µL of sample mixture** and load it onto the sample well on the cartridge. Leave the sample-loaded cartridge at room temperature for **10 minutes.**

FT3

Free Triiodothyronine (FIA)

- Insert the sample-loaded cartridge into the cartridge holder of instrument for infinosiTM tests.
Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder.
- Press “**Test**” button on the instrument for infinosiTM tests.
- Instrument for infinosiTM tests will start scanning the sample-loaded cartridge immediately.
- Read the test result on the display screen of the instrument for infinosiTM tests.
- Print out the testing results when press “**Print**” button on the instrument for infinosiTM tests.

Limitations - interference

- The assay is unaffected by icterus (bilirubin < 600 µmol/L or < 35 mg/dL), hemolysis (Hb < 0.559 mmol/L or < 0.9 g/dL), lipemia (Intralipid < 1200 mg/dL), and biotin < 94 nmol/L or < 23 ng/mL.
- Criterion: Recovery within ± 10 % of initial value.
- Heterophilic antibodies and rheumatoid factors in samples may interfere with test results. Heterophilic antibodies in human serum can react with reagent immunoglobulins, interfering with in vitro immunoassays. Patients routinely exposed to animals or animal serum products can be prone to this interference and anomalous values may be observed. Additional information may be required for diagnosis. This kind of samples is not suitable to be tested by this assay.
- In severe NTI (nonthyroidal illness), the assessment of thyroid status becomes very difficult. TSH measurements are recommended to identify thyroid dysfunction.
- Familial dysalbuminemic conditions may yield erroneous results on direct free T3 assays.
- If a patient, for some reason, reads higher than the highest calibrator report as such (e.g. > 50 pmol/l). Do not try to dilute the sample. TBG variations in different matrices will not allow FT3 hormones to dilute serially.

Measuring range

0.50-50 pmol/L (defined by the lower detection limit and the maximum of the master curve). Values below the detection limit are reported as < 0.50 pmol/L. Values above the measuring range are reported as > 50 pmol/L.

Lower detection limit

0.50 pmol/L

The detection limit represents the lowest analyte level that can be distinguished from zero. It is calculated as the value lying two standard deviations above that of the lowest standard (master calibrator, standard 1+2 SD, repeatability study, n = 21).

Expected values

3.30-7.50 pmol/L

These values correspond to the 2.5th and 97.5th percentiles of results obtained from a total of 127 healthy test subjects examined.

We have not studied the reference intervals in children, adolescents and pregnant women.

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data are given below. Results obtained in individual laboratories may differ.

Precision

Intra-assay

Determined by using 10 tests in the same batch to test with FT3 control, CV ≤ 15%

Inter-assay

Determined by using 3 tests in 3 random and continuous batches to test with FT3 control, CV ≤ 20%

Method comparison

A comparison of the infinosiTM FT3 assay (y) with the Roche Elecsys FT3 (x) using clinical samples gave the following correlation:

Number of samples measured: 105

Linear regression

$y = 1.0347x - 0.2789$

$r = 0.9705$

infinosiTM

Analytical specificity

For the antibody derivative used, the following cross-reactivities were found: D-T3 100 %; L-T4 < 0.31 %; D-T4 < 0.45 %; L-rT3 < 0.05 %; L-T2 < 0.8 %.

Functional sensitivity














0.55 pmol/L

The functional sensitivity is the lowest analyte concentration that can be reproducibly measured with an intermediate precision CV of ≤ 20 %.

References

- Wheeler MH, Lazarus JH. Diseases of the Thyroid. London, Glasgow, Weinheim, New York, Tokyo, Melbourne, Madras: Chapman and Hall Medical, 1994:107-115.
- Pfannenstiel P, Saller B. Schilddrüsenerkrankungen Diagnose und Therapie. Berliner Medizinische Verlagsanstalt GmbH 1995;2:30-32,60-62.
- Fisher DA. Physiological variations in thyroid hormones; physiological and pathophysiological considerations. Clin Chem 1996;42:135-139.
- Bartelena L, Robbins J. Variations in Thyroid Hormone Transport Proteins and Their Clinical Implications. Thyroid. 1992;2(3):237-245.
- Carrero JJ, Qureshi AR, Axelsson J, et al. Clinical and biochemical implications of low thyroid hormone levels (total and free forms) in euthyroid patients with chronic kidney disease. J. Intern. Med. 2007;262(6):690-701.
- Verheeecke P. Free triiodothyronine concentration in serum of 1050 euthyroid children is inversely related to their age. Clin Chem. 1997;43(6):963-967.

Symbols

 In vitro diagnostic medical device	 Temperature limit	 Consult instructions for use	 Catalog number
 Batch code	 Date of manufacture	 Use-by date	 Contains sufficient for <n> tests
 Manufacturer	 Do not re-use	 Do not use if package is damaged and consult instructions for use	 European Conformity
 Authorized representative in the European Community			



diasino

DiaSino Laboratories Co., Ltd
6/F, No.3 Building, No.68 Jingnansi Road National Eco & Tech Development Area
Zhengzhou Henan 450000 China



CMC Medical Devices & Drugs S.L.
C/Horacio Lengo N° 18, CP 29006 Málaga, Spain. Tel: +34 9512 14054